

K051A49

Special 510(k) Premarket Notification
GE Healthcare - GE Vivid 7 and EchoPAC BT05
June 1, 2005

JUN 16 2005

Attachment B:
Summary of Safety and Effectiveness
Prepared in accordance with 21 CFR Part 807.92(c).



GE Medical Systems

General Electric Company
P.O. Box 414, Milwaukee, WI 53201

Section a):

1. **Submitter:** GE Medical Systems, Ultrasound and Primary Care Diagnostics, LLC
PO Box 414, Milwaukee, WI 53201

Contact Person: Allen Schuh,
Manager, Safety and Regulatory Engineering
Telephone: 414-647-4385; Fax: 414-647-4090

Date Prepared: June 1, 2005
2. **Device Name:** GE Vivid 7 Diagnostic Ultrasound System with EchoPAC BT05
Ultrasonic Pulsed Echo Imaging System, 21 CFR 892.1560, 90-IYO
Ultrasonic Pulsed Doppler Imaging System, 21 CFR 892.1550, 90-IYN
Diagnostic Ultrasonic Transducer, 21 CFR 892.1570, 90-ITX
3. **Marketed Device:** GE Vivid 7 Diagnostic Ultrasound System K003931/K041552 (90-IYO/IYN/ITX)
A device currently in commercial distribution.
4. **Device Description:** The GE Vivid 7 Diagnostic Ultrasound is a full featured echocardiography imaging and analysis system. It consists of a mobile console approximately 65 cm wide, 96 cm deep and 139 cm high that provides digital acquisition, processing and display capability. The user interface includes a computer keyboard, specialized controls and a color video CRT display. This modification will provide users with for improved performance and productivity.
5. **Indications for Use:** The device is intended for use by a qualified physician for ultrasound evaluation of Fetal; Abdominal (including renal and GYN); Pediatric; Small Organ (breast, testes, thyroid); Neonatal Cephalic; Adult Cephalic; Cardiac (adult and pediatric); Peripheral Vascular (PV); Musculo-skeletal Conventional; Urology (including prostate), Transesophageal; Transrectal (TR); Transvaginal (TV); and Intraoperative (abdominal, thoracic, & vascular).
6. **Comparison with Predicate Device:** The GE Vivid 7 BT05 is of a comparable type and substantially equivalent to the current GE Vivid 7 with enhanced measurement capability, upgraded beamformer and improved patient exam reporting. It has the same overall characteristics, key safety and effectiveness features, physical design, construction, and materials, and has the same intended uses and operating modes as the predicate device.

Section b):

1. **Non-clinical Tests:** The device has been evaluated for acoustic output, biocompatibility, cleaning and disinfection effectiveness, electromagnetic compatibility, as well as thermal, electrical and mechanical safety, and has been found to conform with applicable medical device safety standards.
2. **Clinical Tests:** None required.
3. **Conclusion:** Intended uses and other key features are consistent with traditional clinical practice, FDA guidelines, and established methods of patient examination. The design and development process of the manufacturer conforms with 21 CFR 820, ISO 9001:2000 and ISO 13485 quality systems. The device conforms to applicable medical device safety standards and compliance is verified through independent evaluation with ongoing factory surveillance. Diagnostic ultrasound has accumulated a long history of safe and effective performance. Therefore, it is the opinion of GE Healthcare that the GE Vivid 7 BT05 Diagnostic Ultrasound is substantially equivalent with respect to safety and effectiveness to devices currently cleared for market.



JUN 16 2005

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Mr. Allen Schuh
Manager, GE Ultrasound Safety and Regulatory Engineering
GE Medical Systems, Ultrasound and Primary Care Diagnostics, LLC
4855 West Electric Avenue
WEST MILWAUKEE WI 53219

Re: K051449
Trade Name: GE Vivid 7 and EchoPAC BT05
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulation Number: 21 CFR 892.1560
Regulation Name: Ultrasonic pulsed echo imaging system
Regulation Number: 21 CFR 892.1570
Regulation Name: Diagnostic ultrasonic transducer
Regulatory Class: II
Product Code: IYN, IYO, and ITX
Dated: June 1, 2005
Received: June 2, 2005

Dear Mr. Schuh:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the GE Vivid 7 and EchoPAC BT05, as described in your premarket notification:

Transducer Model Number

4C
M4S

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This determination of substantial equivalence is granted on the condition that prior to shipping the first device, you submit a postclearance special report. This report should contain complete information, including acoustic output measurements based on production line devices, requested in Appendix G, (enclosed) of the Center's September 30, 1997 "Information for Manufacturers Seeking Marketing Clearance of Diagnostic Ultrasound Systems and Transducers." If the special report is incomplete or contains unacceptable values (e.g., acoustic output greater than approved levels), then the 510(k) clearance may not apply to the production units which as a result may be considered adulterated or misbranded.

The special report should reference the manufacturer's 510(k) number. It should be clearly and prominently marked "ADD-TO-FILE" and should be submitted in duplicate to:

Food and Drug Administration
Center for Devices and Radiological Health
Document Mail Center (HFZ-401)
9200 Corporate Boulevard
Rockville, Maryland 20850


This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>

Page 3 – Mr. Schuh

If you have any questions regarding the content of this letter, please contact Rodrigo C. Perez at (301) 594-1212.

Sincerely yours,

for 

Nancy C. Brogdon
Director, Division of Reproductive,
Abdominal and Radiological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure(s)

Diagnostic Ultrasound Indications for Use Form

GE Vivid 7 Ultrasound System

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	RT3D Mode*
Ophthalmic											
Fetal / Obstetrics	P	P	P	P	P	P	P	P	P	P	P
Abdominal ^[1]	P	P	P	P	P	P	P	P	P	P	P
Pediatric	P	P	P	P	P	P	P	P	P	P	P
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	
Neonatal Cephalic	P	P	P	P	P	P	P	P	P	P	
Adult Cephalic	P	P	P	P	P	P	P	P	P	P	P
Cardiac ^[3]	P	P	P	P	P	P	P	P	P	P	P
Peripheral Vascular	P	P	P	P	P	P	P	P	P	P	
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Superficial											
Other ^[4]	P	P	P	P	P	P	P	P	P	P	P
Exam Type, Means of Access											
Transesophageal	P	P	P	P	P	P	P	P	P	P	
Transrectal	P	P	P		P	P	P	P		P	
Transvaginal	P	P	P		P	P	P	P		P	
Transurethral											
Intraoperative ^[5]	P	P	P		P	P	P	P		P	
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes renal, GYN/Pelvic

[2] Small organ includes breast, testes, thyroid.

[3] Cardiac is Adult and Pediatric.

[4] Other use includes Urology/Prostate

[5] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV).

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

[*] RT3D is Realtime 3D / 4D volume tissue scan acquisition;

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. DeGroom
(Division Sign-Off)
Division of Reproductive, Abdominal,
and Radiological Devices
510(k) Number 2051449

✓ Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form

GE Vivid 7 with 4C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	Other
Ophthalmic											
Fetal / Obstetrics	N	N	N		N	N	N	N	N	N	
Abdominal ^[1]	N	N	N		N	N	N	N	N	N	
Pediatric	N	N	N		N	N	N	N	N	N	
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	N	N	N		N	N	N	N	N	N	
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	N	N	N		N	N	N	N	N	N	
Exam Type, Means of Access											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN;

[4] Other use includes Urology;

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Division Sign-Off)

Division of Reproductive, Abdominal,
 and Radiological Devices

510(k) Number

K051449

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form

GE Vivid 7 with M4S Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics	N	N	N	N	N	N	N	N	N	N	
Abdominal ^[1]	N	N	N	N	N	N	N	N	N	N	
Pediatric	N	N	N	N	N	N	N	N	N	N	
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic	N	N	N	N	N	N	N	N	N	N	
Cardiac ^[3]	N	N	N	N	N	N	N	N	N	N	
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other (specify) ^[4]	N	N	N	N	N	N	N	N	N	N	
Exam Type, Means of Access											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes Renal and GYN;


[3] Cardiac is Adult and Pediatric.

[4] Other use includes Urology;

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number 2051449

✓ Prescription User (Per 21 CFR 801.109)